

GB Gebruiksaanwijzing

Produktbeschreibung

HydroTac Transparent™ is ein transparenter Hydroverband aus hydratisiertem Polyurethan. Die Rückseite besteht aus einer wasserdampfdurchlässigen, flüssigen- und keimdichten Polyurethan-Folie.

HydroTac Transparent Comfort verfügt zusätzlich über einen umlaufenden, selbsthaftenden Kleberand, der mit einem auf Acrylat basierten Kleber beschichtet ist.

Eigenschaften und Wirkungsweise

HydroTac Transparent kann sowohl Feuchtigkeit abgeben als auch in begrenztem Maße Esssud aufnehmen. Hierdurch stellt sich ein feuchtes Wundmilieu ein, das stimulierend auf die Wundheilung wirkt. HydroTac Transparent ist transparent und ermöglicht einen atraumatischen Verbandwechsel.

Durch die Transparenz ist eine Wundinspektion jederzeit und ohne Verbandwechsel möglich. Da HydroTac Transparent Comfort über einen haarfrendlichen umlaufenden Kleberand verfügt, sind keine zusätzlichen Materialien zur Verbandfixierung notwendig.

Zweckbestimmung/Indikationen

HydroTac Transparent eignet sich für die Behandlung von trockenen Wunden oder Wunden mit geringer Exsudatmenge (in der Granulations- und Epithelisierungsphase).

HydroTac Transparent sollte von medizinschen Fachkräften verwendet werden. Diese können den Verbandwechsel an die Patienten und/oder ihre Angehörigen (Laien unter Anleitung von den Fachkräften) delegieren.

Anwendungsgebiete
HydroTac Transparent ist der Größe so auszuwählen, dass die Wundkante die Wundränder ca. 1 – 2 cm überlappt. Falls erforderlich, kann HydroTac Transparent auf die gewünschte Größe zugeschnitten werden. Dafür nur eine sterile Schere verwenden. Die umliegende Haut sollte trocken sein, bevor HydroTac Transparent Comfort mit umlaufendem Kleberand angebracht wird.

HydroTac Transparent: Die blaue Schutzfolie entsprechend der Nummerierung (1 & 2) abziehen. Den Verband mit der geschützten Seite auf die Wunde legen und mit einem geeigneten Sekundärverband fixieren.

HydroTac Transparent Comfort: Die blaue Schutzfolie entsprechend der Nummerierung (1 & 2) abziehen, den Verband mit der geschützten Seite auf die Wunde legen und den Kleberand gut andrücken. Dann die schützende Deckfolie entsprechend der Nummerierung (3 & 4) abziehen.

HydroTac Transparent kann bis zu 7 Tage auf der Wunde verbleiben. HydroTac Transparent zum Verbandwechsel an einer Ecke leicht lösen und vorsichtig abziehen.

Produktinspektion
Um das Risiko potentieller Infektionsfahen oder einer Umweltschutzung zu minimieren, sollten die WevegerKomponenten von HydroTac Transparent gemäß geltender lokaler Gesetzen, Verordnungen, Vorschriften und Standards zur Infektionsprävention entsorgt werden.

Kontraindikationen
HydroTac Transparent nicht bei Patienten mit einer Unverträglichkeit gegen einen der Bestandteile, insbesondere Polyurethan, Glyzerin oder einen auf Acrylat basierten Kleber, anwenden. Nicht anwenden, wenn eine klinische Infektion vorliegt, desgleichen bei Brandwunden 3. Grades.

Mögliche Nebenwirkungen
Wirdes andere Medikation/Verträglichkeit kann die Wundauflage Nebenwirkungen haben, die nicht bei jedem Patienten auftreten.

Allgemeine Nebenwirkungen von HydroTac Transparent:
• In seltenen Fällen kann HydroTac Transparent eine lokale Reaktion hervorrufen (z.B. Juckreiz, Schwellung oder Rötung).
• In seltenen Fällen kann HydroTac Transparent eine allergische Reaktion hervorrufen

Wenn Nebenwirkungen bei Patienten auftreten, auch wenn die Hier nicht aufgeführt sind, wenden Sie sich bitte an Ihren Arzt oder Apotheker.

Besondere Vorsichtsmaßnahmen
Vor jeder Behandlung von Wunden mit beeinträchtigter Heilungstendenz ist eine ärztliche Beurteilung des Wundzustandes und der Ursachen aller Wundheilungsstörungen notwendig. Die Behandlung mit HydroTac Transparent kann eine Kausalbehandlung der Wundheilungsstörungen nicht ersetzen. Die Wiederanfertigung von Produkten für den Einmalgebrauch weggehabene Medizinpakete ist gefährlich. Die Wiederaufbereitung von Einmalprodukten, um sie erneut zu verwenden, kann ihre Integrität und Leistung deutlich beeinträchtigen.

Vom Öffnen die Produktverpackung sorgfältig auf Beschädigungen überprüfen. Das Produkt nicht verwenden, wenn sichtbare Abzehrungen, Abzehrungen oder unvollständige Siegelung, Einstülplicher oder ungleichmäßige, abgeblöste, gerissene oder unvollständige Siegelung.

Meldung von Vorkommnissen
Für Patienten/Anwender/Dritte in der Europäischen Union und in Ländern mit denselben Rechtsvorschriften (Niederlande (UE) 2017/745) oder in Ländern mit anderen Rechtsvorschriften (USA) Verwendung dieses Produktes einschließlich des vorgeschriebenen Vorkommis aufgetreten ist, melden Dies dem Hersteller und/oder seinen Bevollmächtigten und Ihrer nationalen Behörde.

Allgemeine Hinweise
Bei längerer Lagerung kann eine leichte Gelbfärbung auftreten, welche keinerlei Auswirkungen auf die Produkteigenschaften hat. Für Kinder ungeeignet aufbewahren.

HydroTac Transparent im folgenden Text angedeutet wird, bezieht sich dies auf alle Aufmachungen, einschließlich der Comfort-Varianten; davon abweichende Ausnahmen sind ausdrücklich erwähnt.

Stand der Information: 2024-04-07
DE – PAUL HARTMANN AG - 89522 HEIDENHEIM, GERMANY
AT – PAUL HARTMANN Ges.m.b.H. - 2355 Wiener Neudorf
CH – IFF HARTMANN AG - 8212 Neuhausen

GB Instructions for use
HydroTac Transparent™ is a transparent hydrogel dressing consisting of hydrated polyurethane. The backing consists of a water vapour permeable polyurethane film which is impermeable to liquids and bacteria.

In addition, HydroTac Transparent Comfort has an adhesive border, which is coated with an acrylic based adhesive.

Properties and mode of action
HydroTac Transparent can release moisture and absorb a limited amount of exudate. This provides a moist wound environment which has a stimulating effect on wound healing. HydroTac Transparent is transparent and provides an atraumatic dressing change. Because of the transparency, wound inspection is possible at any time without having to remove the film border.

As HydroTac Transparent Comfort has a skin-friendly adhesive dressing border, no other materials are needed to secure the dressing.

Intended use / Indications
HydroTac Transparent™ is suitable for the treatment of dry or lightly exuding wounds in the granulation and epithelialization stages.

HydroTac Transparent is intended to be applied by healthcare professionals. These may delegate dressing changes to patients and/or their relatives (non-professionals under guidance of professionals).

Mode of application
Select a size of HydroTac Transparent to suit the wound pad overlaps the margins of the wound by approximately 1 – 2 cm. If required, the product HydroTac Transparent may be cut to desired size, using sterile scissors only. Surrounding skin should be dry before applying HydroTac Transparent Comfort with adhesive border.

HydroTac Transparent: Remove the blue protective layer following the numbering (1 & 2). Place the dressing with the gel-impregnated side on the wound and fix with suitable secondary dressing material.

HydroTac Transparent Comfort: Remove the blue protective layer following the numbering (1 & 2), place the dressing with the gel-impregnated side on the wound and press down the adhesive edges firmly. Remove the protective top layer following the numbering (3 & 4).

HydroTac Transparent can be left on wounds up to seven days. To change the dressing, detach HydroTac Transparent slightly at one edge and remove it carefully.

Product disposal
To minimize the risk of potential infection hazards or environmental pollution, disposable components of HydroTac Transparent should follow disposal procedures according to applicable and local laws, rules, regulations and infection prevention standards.

Contraindications
Do not use HydroTac Transparent on patients with an intolerance to any of its ingredients, especially polyurethane, glycerin or acrylic based adhesive. Do not use on wounds which are clinically infected or on 3rd degree burns.

Possible adverse events
As any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

As with any other medical device, this product can have side effects, which do not necessarily occur for everybody.

General side effects of HydroTac Transparent:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

HydroTac Transparent Comfort:
• In rare cases HydroTac Transparent might cause local reactions (irritation, skin reactions e.g. redness, swelling).

• In rare cases HydroTac Transparent might cause allergic reactions

Segnalazione di incidenti

Per un paziente/utente/terza parte nell'Unione europea e in paesi con regime normativo identico (Regolamento (UE) 2017/745 sui dispositivi medici), se, durante l'uso del presente dispositivo o come risultato del suo utilizzo, si è verificato un incidente grave, segnalare quest'ultimo al fabbricante o al rappresentante autorizzato e alla propria autorità nazionale.

Informazioni generali

*Questo prodotto può scolorire nel tempo senza impatto sulle prestazioni del prodotto. Tenere lontano dalla portata dei bambini.

*Prodotti riferimenti a HydroTac Transparent presente nel testo si riferiscono a tutte le confezioni, comprese le versioni Comfort. Le eccezioni dei diffezionati da queste sono esplicitamente menzionate.

Data di revisione del testo: 2024-04-07
IT – PAUL HARTMANN S.p.A. - 37139 Verona

ES Instrucciones de uso

Descripción del producto
HydroTac Transparent™ es un apósito de hidrogel transparente a base de poliuretano hidratado. La parte impermeable a líquidos y bacterias.

Adicionalmente, HydroTac Transparent Comfort cuenta con un rebordo de fijación adhesivo, revestido con adhesivo acrílico.

Propiedades y modo de acción
HydroTac Transparent libera humedad y absorbe alguna cantidad de exudado. Este modo, crea un ambiente húmedo, que estimula la cicatrización de la herida. HydroTac Transparent es transparente y garantiza una mudanza de perfo atráumtica. Debido a la transparencia, permite a inspección de la ferida en cualquier altura senza remoción de pelo.

HydroTac Transparent Comfort contém um rebordo adesivo bem tolerado pela pele, não sendo necessário qualquer material extra para fixação do peso.

Utilização prevista/Indicações
HydroTac Transparent é adequado para o tratamento de feridas secas ou com exudado moderado, na fase de granulação e de epitelização.

HydroTac Transparent destina-se à aplicação por profissionais de saúde. Estes deverão delegar a mudança do penso aos pacientes/e aos familiares (não-profissionais sob orientação de profissionais de saúde).

Modo de utilização
Este produto não deve ser utilizado em feridas abertas e/ou com exudado moderado. Este produto não deve ser utilizado em feridas abertas e/ou com exudado moderado. Este produto não deve ser utilizado em feridas abertas e/ou com exudado moderado.

Características y principio de acción
HydroTac Transparent permite liberar humedad y absorber una cantidad limitada de exudado. De este modo, proporciona un medio húmedo que tiene un efecto estimulador en el proceso de cicatrización de la herida. HydroTac Transparent es transparente y permite cambiar el apósito sin dolor.

Si es transparente, permite inspeccionar la herida en cualquier momento sin necesidad de retirar el apósito.

HydroTac Transparent Comfort tiene un borde con un adhesivo repositivo con la piel, por lo que no se necesita nada más para fijar el apósito con firmeza.

Use previsto /Indicações
HydroTac Transparent é adequado para o tratamento de feridas secas ou com poco exudado, em las fases de granulação y epitelização.

HydroTac Transparent está previsto para a aplicação por parte de profissionais sanitários. Estes poderão delegar o cambio de sio em los pacientes y/o familiares (las personas no sanitarias) desberán hacerlo con las indicaciones de expertos).

Modo de aplicación
Elija un HydroTac Transparent del tamaño adecuado para que la almohadilla sobresalga de los bordes de la herida entre 1 y 2 cm. Si fuera necesario, puede recortar el apósito HydroTac Transparent al tamaño deseado, usando únicamente tijeras estériles para este fin. La piel circundante debe estar seca antes de aplicar el apósito HydroTac Transparent Comfort con borde adhesivo.

HydroTac Transparent: retire la lámina protectora azul en el orden que indican los números (1 y 2). Coloque el apósito con el lado impregnado en gel sobre la herida y presiónese con firmes los bordes adhesivos. Retire la lámina protectora de la parte superior en el orden que indican los números (3 y 4).

HydroTac Transparent puede dejar en la herida hasta diez días. Para cambiar el apósito levante el HydroTac Transparent ligemente por una esquina y retírela con cuidado.

HydroTac Transparent: retire la lamina protectora azul en el orden que indican los números (1 y 2), coloque el apósito con el lado impregnado en gel sobre la herida y presiónese con firmes los bordes adhesivos. Retire la lamina protectora de la parte superior en el orden que indican los números (3 y 4).

HydroTac Transparent puede dejar en la herida hasta diez días. Para cambiar el apósito levante el HydroTac Transparent ligemente por una esquina y retírela con cuidado.

HydroTac Transparent: retire la lamina protectora azul en el orden que indican los números (1 y 2), coloque el apósito con el lado impregnado en gel sobre la herida y presiónese con firmes los bordes adhesivos. Retire la lamina protectora de la parte superior en el orden que indican los números (3 y 4).

HydroTac Transparent puede dejar en la herida hasta diez días. Para cambiar el apósito levante el HydroTac Transparent ligemente por una esquina y retírela con cuidado.

HydroTac Transparent: retire la lamina protectora azul en el orden que indican los números (1 y 2), coloque el apósito con el lado impregnado en gel sobre la herida y presiónese con firmes los bordes adhesivos. Retire la lamina protectora de la parte superior en el orden que indican los números (3 y 4).

HydroTac Transparent puede dejar en la herida hasta diez días. Para cambiar el apósito levante el HydroTac Transparent ligemente por una esquina y retírela con cuidado.

HydroTac Transparent: retire la lamina protectora azul en el orden que indican los números (1 y 2), coloque el apósito con el

